



Republic of the Philippines  
**Benguet State University**

2601 La Trinidad, Benguet  
www.bsu.edu.ph

Telefax No. (074) 661-1839; Email:procurement@bsu.edu.ph

**REQUEST FOR QUOTATION (RFQ)**  
**(ALL ENTRIES ARE MANDATORY)**

Name of Owner/ Corporation/ Cooperative/ Agency: \_\_\_\_\_

QUOTATION NO. 1269

Name of Business : \_\_\_\_\_

DATE: NOV 03 2022

Address: \_\_\_\_\_

P.R. NO. : 2022-05-512

Email address: \_\_\_\_\_ Telephone No. \_\_\_\_\_

ABC: Php. 533,000.00

FUND SOURCE: 911 (DA-BAR Project)

The Benguet State University through its Bids and Awards Committee will undertake NP-1VP  
for the item/s listed below under the following terms and conditions:

- All entries must be written legibly and accurately.
- Delivery period is within 30 calendar days.
- Warranty period shall be for a minimum of three months, in case of expendable supplies, and a minimum of one year in case of non- expendable supplies from date of acceptance by the University.
- Price validity shall be for a period of 30 calendar days.
- Compliance with specifications: Kindly specify in the column provided below whether the offer fully complies with the specifications or a counter offer.
- Deadline of submission of quotation is on or before 5:00 PM of Nov. 9, 2022. **Offers may be submitted in sealed or open quotation in person, or through facsimile (074-661-1839) or email (procurement@bsu.edu.ph)**
- The item/s shall be procured by:    a) line item;    b) lot;    c) sub-lot
- Business Permit Number : \_\_\_\_\_, date of issue: \_\_\_\_\_, place of issue: \_\_\_\_\_, (If renewal of Business Permit is still on process, please indicate OR No: \_\_\_\_\_, date of OR : \_\_\_\_\_ for the payment of renewal).
- \* PhilGEPS Registration Number: \_\_\_\_\_

**REYNANTE B. BASCO**  
Chairperson, BAC (Goods & Services)

College/Dept./Office: **HERRC (EPN Project)**

Item No.	Qty	Unit	Item	Technical Specifications (Brand and Model)		ABC	Unit Cost	Total Cost
				COMPLY	COUNTER OFFER			
	1	unit	<b>Autoclave</b> Gross Capacity: at least 41 qt / 39 liter Overall Height: at least 19" / 48.3cm Bottom Height: at least 14 1/4" / 36.2cm; Inside Diameter: at least 15 1/4" / 38.7cm; Unit Weight: at least 45 lbs. / 20.4kg.; Inner Container No. 4156 Height: at least 45 lbs. / 20.4kg. Diameter: at least 14" / 35.6cm Circumference: at least 14" / 35.6cm Capacity: at least 27.3 qt / 25.8 liter Volume: at least 1578in3 / 25,856cm3 Carton Dimensions: at least 24" x 24" x 21" 61cm x 61cm x 53.3cm; Cube: 7; Optional No. at least 4180 Support Base: at least 3" / 7.6cm high Outside Diameter: at least 16 1/4" / 41.3cm; Inside Diameter: at least 15" / 38.1cm; Elevates Sterilizer Above Surface: at least 2 1/4" / 5.7cm Geared steam gauge: Equipped with 2 automatic venting features; Pressure regulator weight with settings of 5, 10 and 15 psi; Double thickness edges			70,000.00		
	1	unit	<b>Blender</b> - 10 speed Capacity (l):1.8-2.1L Blending Arm Material:Plastic Voltage (V):220-240V Container Material:Plastic Blade Number:6 Housing Material:Plastic Rotation Rate(RPM):45000 Chopper Material:ABS Number of Speed Settings:Stepless Adjustment Power (W):>1500W			20,000.00		

Item No.	Qty	Unit	Item	Technical Specifications (Brand and Model)		ABC	Unit Cost	Total Cost
				COMPLY	COUNTER OFFER			
	1	unit	<b>Magnetic stirrer/ hot plate:</b> Temperature range: up to 380°C; Speed range: 100 to 1,500 rpm; Top plate material: Ceramic; Top plate size: at least 7 x 7 in. (180 x 180 mm); Dimensions: at least 7.9 x 12.2 x 4.2 in. (20 x 30 x 11.5 cm); Capacity: up to 5L; Watts: 695W; FEATURES: - Precise control of temperature and stirring speeds			28,000.00		
	1	unit	<b>Laminar flow hood/ BOD cabinet:</b> SPECIFICATIONS: Internal size: at least 1200*645*610 mm External size: at least 1300*750*2040 mm; Work surface height: at least 750 mm ; Display: LCD display; Airflow velocity: Average of 0.3 m/s to 0.5 m/s Work zone material: Stainless steel 304 Exterior material: Cold-rolled steel with anti-bacteria powder coating Front and side window: at least 5 mm toughened glass, anti-UV; Pre-filter: Polyester fiber, washable HEPA filter: 99.999% efficiency at 0.3 um;l Noise: < 65 dB; Front window: Motorized operations; Max opening: at least 430 mm; UV lamp: at least 30W*1; emission of 253.7 nm Illumination: at least 12W*1 LED Power consumption: 400W Standard accessories: LED lamp, UV lamp*2, Base stand, Gas Tap, Waterproof socket*2 Optional accessories: Electric height adjustable base stand Caster: Universal wheel with leveling feet; Electricals: AC220V±10%, 50/60Hz ; inclusion of AVR			180,000.00		
	1	unit	<b>Trinocular Compound microscope:</b> SPECIFICATIONS: Optical system: Infinite corrected; Viewing head: Trinocular Head, Inclined at 30°, Interpupillary 47-78mm, 360° Rotation; Nosepiece: Quadruple revolving nosepiece; Eyepiece: Wide Field Eyepiece WF10X/20; with ocular micrometer Objectives: Achromatic objectives: at least 4x/0.10, 10x/0.25, 40x/0.65, spring loaded, 100x/1.25, spring loaded, oil Stage: Rackless Double Layer Mechanical Stage at least 150X139mm, Moving Range at least 75X52mm; Focusing: Coaxial Coarse and Fine Adjustment, Fine Division at least 0.002mm, Coarse Stroke at least 37.7mm per Rotation, Fine Stroke at least 0.2mm per Rotation, Moving Range at least 25mm; Condenser: NA 1.25 Abbe Condenser with Aperture Diaphragm; Illumination: LED illumination for long life, brightness adjustable; Supplied with: Dust cover, spare fuse, blue, green filters; Power: 240V, 50/60Hz; Inclusion: HDMI & WIFI DIGITAL CAMERA, 2MP/5MP  Sensor & Size(mm): 1080P/1.8" (6.22x4.67); Pixel(µm): 2.4x2.4 G Sensitivity: at least 425mv with 1/30s Dark Signal: at least 0.15mv with 1/30s FPS/Resolution: at least 30@1920x1080 (HDMI), 25@1920x1080 (WiFi); Binning: 1x1; Exposure: 0.03ms~918ms, inclusion of AVR			120,000.00		

Item No.	Qty	Unit	Item	Technical Specifications (Brand and Model)		ABC	Unit Cost	Total Cost
				COMPLY	COUNTER OFFER			
	1	unit	<b>Stereomicroscope:</b> PECIFICATIONS: Viewing Heads: Trinocular, 360° rotating and 45° inclined; Interpupillary distance: Adjustable between 54 and 76 mm Eyepieces: WF10x/20 mm Specimen stage: Fitted with a transparent disc for transmitted light with specimen clips; Objectives: Continuous zoom 0.7x – 4.5x. (6.43:1 zoom factor); Working distance: 85 mm; Stand: Fixed arm type with focusing mechanism; Focusing: Rack and opinion controlled by a pair of knobs placed on both sides of the stand; Illumination: Double adjustable illuminator for incident and transmitted light fitted with two high-efficiency LED light sources; the angle of the incident illuminator is adjustable; Power: 100-240V, 50/60Hz; Inclusion: ICOE HUWS5, HUWS SERIES HDMI & WIFI DIGITAL CAMERA, 2MP/5MP; Sensor & Size(mm): 1080P/5M/Sony IMX178@ 1/1.8" (6.22x4.67); Pixel(µm): 2.4x2.4 G Sensitivity: at least 425mv with 1/30s Dark Signal: at least 0.15mv with 1/30s FPS/Resolution: at least 30@1920x1080 (HDMI), 25@1920x1080 (WIFI) ; Binning: at least 1x1; Exposure: at least 0.03ms-918ms; inclusion of AVR			90,000.00		
	1	unit	<b>Vortex mixer. Speed:</b> 200 to 3200 rpm; Operating modes: Touch or continuous; Orbit: 3 mm; Operating Temp. +4 to +45°C; Dimensions: (w x d x h) 5 x 6.3 x 6.7 in. 13 x 16 x 17 cm; Electrical: 220V, 50Hz, 180W; inclusion of AVR			25,000.00		
<b>TOTAL</b>								
<b>XXXXXXXXXXXX</b>								

Counter offers:

Delivery Period: \_\_\_\_\_  
 Warranty: \_\_\_\_\_  
 Price Validity: \_\_\_\_\_

We hereby submit our quotations, accept the general conditions for the above items and be bound by our offer.

\_\_\_\_\_  
Signature over Printed Name

\_\_\_\_\_  
Date and Time of Receipt of RFQ

\* Not a ground for disqualification for failure to indicate but required as a condition for award.

Revised: 9/19/2022

12th 512